			<b>'</b>
	Application No.	Applicant(s)	
Notice of Allowability	09/957,461	MACHIDA, YOSHIO	
	Examiner	Art Unit	-
	Julianne M. Sullivan	3768	
The MAILING DATE of this communication appear All claims being allowable, PROSECUTION ON THE MERITS IS herewith (or previously mailed), a Notice of Allowance (PTOL-85) NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT R	(OR REMAINS) CLOSED in this ap or other appropriate communication IGHTS. This application is subject t	oplication. If not include n will be mailed in due	ed course. THIS
1. $\boxtimes$ This communication is responsive to <u>the Request for Conti</u>	inued Examination filed January 26,	<u>2006</u> .	
2. X The allowed claim(s) is/are 1,3-14,16-20 and 22-26.			
3.			
Attachment(s)  1. Notice of References Cited (PTO-892)  2. Notice of Draftperson's Patent Drawing Review (PTO-948)  3. Information Disclosure Statements (PTO-1449 or PTO/SB/0 Paper No./Mail Date  4. Examiner's Comment Regarding Requirement for Deposit of Biological Material	8. 🛭 Examiner's Statem	y (PTO-413), ate Iment/Comment	
	9.  Other		

## **EXAMINER'S AMENDMENT**

1. An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

Authorization for this examiner's amendment was given in a telephone interview with Larry S. Nixon on April 17, 2006.

The application has been amended as follows:

In Claim 14, line 2: delete "trhee" and insert --three--;

In Claim 14, line 14: between "as" and "synchronously" insert --to--.

## **REASONS FOR ALLOWANCE**

2. The following is an examiner's statement of reasons for allowance: the prior art does not teach or suggest an MRI system where three or more slices are excited in an alternating and interleaved pattern while the object being imaged is continuously moved, and where the three or more imaging regions synchronously track the movement of the object. Additionally, the present invention is distinguished from the prior art in that plural slices are not excited one at a time as they move through the imaging volume, but are repeatedly excited using the alternating and interleaved technique such that multiple slices are imaged in the imaging volume at the same time.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue

Art Unit: 3768

fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Julianne M. Sullivan whose telephone number is 571-272-6084. The examiner can normally be reached on Monday through Friday 8:00am to 4:30pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eleni Mantis Mercader can be reached on 571-272-4740. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

BRIAN L. CASLER
SUPERVISORY PATENT EXAMINER
TECHNOLOGY CENTER 3700